湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

地址(Add): 湖北省武汉市东湖新技术开发区茅 店山中路二号

网址 (Web site): http://www.hbjl.gov.cn

邮编 (Post Code): 430223 电话 (Tel): 027-81925136

传真 (Fax): 027-81925137

中国校准服务联合体 China United Calibration Service



REGISTRATION NO. 002

CALIBRATION CERTIFICATE

证书编号:

[2019DW03540776]

Certificate No.

委托方名称

中广核太阳能德令哈有限公司

Customer 委托方地址

青海省海西州德令哈市西出口315国道以北

Address

样品名称

Name Of Sample

蓄电池放电仪

制造厂商

武汉国电华美电气设备有限公司

Manufacturer

HM3982-220V200A

型号规格

Model/Type

器具编号 No Of Sample

1911071

证书专用章 Stamp






校准日期 Calibration date 2019 年 12 月 12 日 M

证书批准人 Approved by

核验员 Checked by

校准员 Calibrated by



本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研 究院于1999年代表中国签署了国际间"国家计量基标准及国家计量研究院出具的校准和测量证书 相互承认协议"。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

证书编号: [2019DW03540776]号

Certificate No.

本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

本次校准的技术依据(名称、代号)

Reference documents for the Calibration (Name 、 Code)

参照: JJF 1587-2016 数字多用表校准规范

本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

数字多用表

型号/编号

Model/Serial No.

8508A/974056743

证书号/有效期

Certificate No./Due Date

DLsc2019-2059/2020-09-25

校准环境条件

Environmental condition on the Calibration

温度:

20. 7°C

相对湿度: 45 %

其它: ____

Temperature

R.H.

Others

气压: 一

地点: 本院光谷基地B102

原始记录编号: 2019DW03540776

Pressure

Place

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated. 未经本院许可,不得部分复制本证书。

证书编号: [2019DW03540776]号

Certificate No.

校准数据/结果

Data/Results of Calibration

一、电压测量		
标准值/V	被检显示值/V	测量不确定度(<i>U</i> ; <i>k</i> =2)
50.0	50.1	0.2V
100.0	99.5	0.2V
150.0	149.0	0.3V
200.0	198.3	0.3V
220.0	218.0	0.3V
260.0	257.6	0.3V

以下空白

